

QUESTION & ANSWERS N°1- dated 10.03.2022

Request for Offer 06/2022

related to

LOT 2: Supply of water quality monitoring equipment and accessories in Peja

Q1	You have requested for item 6 - Magnetic stirring bars Pack (10 pieces). - Please confirm we can offer Magnetic stirring bars cylindrical with length: 25 mm and diameter: 6 mm?
A1	We confirm the following: The 'Length 30 mm and diameter 6 mm' is preferable, however the proposed 'Length 25 mm and diameter 6 mm' is acceptable too
Q2	You have requested for item 8 - Water distiller system 5L/h - Please confirm we can offer Water distiller system with production output: 4 L/h and tank volume: 8 L or Water distiller system with production output: 8 L/h and tank volume: 16 L?
A2	We confirm the following: Water distiller system with production output: 4 L/h and tank volume: 8 L
Q3	You have requested for item 16 – Micropipettes, 1, 5, 10 ml, quantity 3 piece. - Does this mean 3 pipettes – 1 piece for each size, or 3 pipettes for each size?
A3	We confirm the following: 1 piece for each size
Q4	You have requested for item 21 - Pipette tips, 1, 5, 10 ml (1 pack for each) and quantity 3 piece. - Could you please clarify how many tips per pack?
A4	We confirm the following: 100 tips per pack
Q5	You have requested for item 24 - Test tubes – glass, 16 mm. - Please confirm us if we should offer test tubes - glass size 16 x 100mm or 16 x 160mm?
A5	We confirm the following: Test tubes size 16 x 100mm
Q6	You have requested for item 29 - Imhoff cones for sediment. - Please confirm if we should offer Imhoff cones with or without stopcock?
A6	We confirm the following: Imhoff cones without stopcock
Q7	You have requested for item 32 - Filter paper, 120 mm and quantity 5 piece. - Since the standard diameters of Filter papers are 110 mm and 125mm, please confirm that we should offer Filter paper with 125 mm diameter?

Description:

Dimensions: x