

QUESTION & ANSWERS N°1- dated 10.03.2022

Request for Offer 06/2022

related to

LOT 2: Supply of water quality monitoring equipment and accessories in Peja

Q1	You have requested for item 6 - Magnetic stirring bars Pack (10 pieces). - Please confirm we can offer Magnetic stirring bars cylindrical with length: 25 mm and diameter: 6 mm?
A1	We confirm the following: The 'Length 30 mm and diameter 6 mm' is preferable, however the proposed 'Length 25 mm and diameter 6 mm' is acceptable too
Q2	You have requested for item 8 - Water distiller system 5L/h - Please confirm we can offer Water distiller system with production output: 4 L/h and tank volume: 8 L or Water distiller system with production output: 8 L/h and tank volume: 16 L?
A2	We confirm the following: Water distiller system with production output: 4 L/h and tank volume: 8 L
Q3	You have requested for item 16 – Micropipettes, 1, 5, 10 ml, quantity 3 piece. - Does this mean 3 pipettes – 1 piece for each size, or 3 pipettes for each size?
A3	We confirm the following: 1 piece for each size
Q4	You have requested for item 21 - Pipette tips, 1, 5, 10 ml (1 pack for each) and quantity 3 piece. - Could you please clarify how many tips per pack?
A4	We confirm the following: 100 tips per pack
Q5	You have requested for item 24 - Test tubes – glass, 16 mm. - Please confirm us if we should offer test tubes - glass size 16 x 100mm or 16 x 160mm?
A5	We confirm the following: Test tubes size 16 x 100mm
Q6	You have requested for item 29 - Imhoff cones for sediment. - Please confirm if we should offer Imhoff cones with or without stopcock?
A6	We confirm the following: Imhoff cones without stopcock
Q7	You have requested for item 32 - Filter paper, 120 mm and quantity 5 piece. - Since the standard diameters of Filter papers are 110 mm and 125mm, please confirm that we should offer Filter paper with 125 mm diameter?

	- Also please confirm if we should offer with color code blue, white or black?
A7	We confirm the following: Filter papers 125 mm, color code: white
Q8	You have requested for item 34 - Bottles for sampling – plastic, 1000, 500 ml and quantity 50 piece. - Please confirm if we should offer 50 pieces for each volume – in total 100 pcs or in total 50 pieces - 25 pieces for volume 1000 ml and 25 pieces for volume 500ml?
A8	We confirm the following: 25 pieces for volume 1000 ml and 25 pieces for volume 500ml
Q9	You have requested for item 36 - Beakers – Glass, 1000, 500, 250, 100, 50 and quantity 25 piece. - Please confirm if we should offer 25 pieces in total of beakers (with requested volumes) or we should offer 25 pieces of each volume of beaker so in total 125 pieces?
A9	We confirm the following: 5 pieces from each/ 25 pieces total
Q10	You have requested for item 37 – Filters, 0.45 mikron and quantity 1 piece. - Since the standard packages of membrane filters are 1x100 pieces, please confirm we should we offer package of 1x100 pcs?
A10	We confirm the following: 1x100 pcs
Q11	You have requested for item 39 – Goggles, Medium, Large. - Please confirm we can offer one size Goggles which can be adapted to individual face curvature providing complete side protection?
A11	We confirm the following: Simple lab safety goggles, one size or other